

*VECTOR*



# **Product list**

## **2025**



## Our contacts

### VECTOR-BEST

PO BOX 492, Novosibirsk, Russia, 630117

Tel: +7 (383) 25-25-163

E-mail: [vbmarket@vector-best.ru](mailto:vbmarket@vector-best.ru)

[www.vector-best.com](http://www.vector-best.com)

## Our distributors

Germany

### Immundiagnostik AG

Stubenwald-Allee 8a

64625 Bensheim

Germany

Tel: +49 6251 70190 0

Fax: +49 6251 70190 363

E-mail: [info@immundiagnostik.com](mailto:info@immundiagnostik.com)

India

### INBIOS

D-10/5, 202, 2nd Floor, Okhla Industrial Area, Phase-II, New Delhi, India  
(110020)

Tel: +91 9821193636

E-mail: [info@inbiosindia.com](mailto:info@inbiosindia.com)

[www.inbiosindia.com](http://www.inbiosindia.com)

### INBIOSYS SYSTEMS PVT.LTD

HOUSE NO.-48, 2ND FLOOR

BLOCK -D, POCKET -1, SEC-16

ROHINI, DELHI -110089

Tel: +91 9811055995

E-mail: [paathakamit@yahoo.com](mailto:paathakamit@yahoo.com)

[www.inbiosyssystems.com](http://www.inbiosyssystems.com)

### IMMUNOCONCEPT INDIA Pvt. Ltd.

426, Patparganj, Industrial Area

NEW DELHI-110092.

Tel: +011 43063564,42547608

Tel: +011 43063564,42547608

E-Mail: [helpdesk@immunoconceptindia.co](mailto:helpdesk@immunoconceptindia.co)

[sales@immunoconceptindia.co](mailto:sales@immunoconceptindia.co)

[Harendra@immunoconceptindia.co](mailto:Harendra@immunoconceptindia.co)

[www.immunoconceptindia.co](http://www.immunoconceptindia.co)

Republic of  
North  
Macedonia

**AVICENA DOO**  
ul. 1551 br.5B, Ind.zona Vizbegovo  
1000 Skopje,  
Republic of North Macedonia  
Tel: +389 2 3079 700  
Fax.: +389 2 3079 400  
E-mail: [avicena@avicena.com.mk](mailto:avicena@avicena.com.mk)  
[www.avicena.com.mk](http://www.avicena.com.mk)

Slovakia

**Diagnostické centrum DNK, s.r.o.**  
Brestova 14, 821 02 Bratislava, Slovakia, EU  
Tel: +421 903 822 515  
E-mail: [dnk@pharma.sk](mailto:dnk@pharma.sk)  
[www.pcr.sk](http://www.pcr.sk)

Greece

**INTERLAB Ltd**  
163-165 P. Tsaldari str., 176 76 Kallithea - Athens, GREECE  
Tel: +30 210 953 2220  
Mob. tel: +30 69 44 54 17 57  
Fax: +30 210 953 2221  
E-mail: [interlab@tee.gr](mailto:interlab@tee.gr)

Serbia

**VBRS D.O.O.**  
Partizanske avijacije 8, 11107 Belgrade, Serbia  
Tel: (+381) 60 062 8234  
E-mail: [petrovic.irena@vbrs.rs](mailto:petrovic.irena@vbrs.rs)

**Product list of the ELISA diagnostic kits**  
**Compliance with the requirements of international standard EN ISO 13485:2016**

**Contents**

Human immunodeficiency virus .....	6
Hepatitis viruses A, B, C, D, E .....	6
Sexually transmitted infections (STIs).....	8
Human herpes viruses and TORCH-infections .....	9
Mycoses .....	11
Tick-borne and zoonotic infections .....	11
Vaccine-preventable diseases.....	13
Respiratory infections.....	13
Tuberculosis .....	14
Parasitic infections .....	14
Gastrointestinal diseases .....	15
Hormones.....	16
Diabetes .....	18
Pregnancy.....	18
Anemia .....	18
Tumor markers.....	19
Acute phase proteins .....	20
Thrombosis.....	20
Cardiac markers .....	20
Renal function .....	21
Systemic and autoimmune diseases.....	21
Humoral immune status .....	22
Cytokines.....	23

## New kits

3980	DHEA-S-EIA-BEST	.....	17
3982	17-OH-progesterone-EIA-BEST	.....	17
8437	CYFRA 21-1-EIA-BEST	.....	19

## Human immunodeficiency virus

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
0134	HIV-1 p24-antigen-EIA-BEST	12×8	Blood serum/plasma	Detection and confirmation of the presence of HIV-1 p24 antigen
0151	CombiBest HIV-1,2 Ag/Ab Set 1	24×8	Blood serum/plasma	Detection of anti-HIV-1, 2 antibodies and HIV-1 p24 antigen
0152	CombiBest HIV-1,2 Ag/Ab Set 2	12×8	Blood serum/plasma	
0193	UniBest HIV-1,2 Ab Set 1/strip	24×8	Blood serum/plasma	Detection of total antibodies to HIV-1, 2
0192	UniBest HIV-1,2 Ab Set 2	12×8	Blood serum/plasma	Two-step analysis
0160	LIA-Blot HIV-1,2 BEST	24	Blood serum/plasma	Confirmation of the presence of antibodies to HIV-1, 2 antigens by immunoblot

## Hepatitis viruses A, B, C, D, E

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
<b>Hepatitis A</b>				
0352	Vectohep A-IgM	12×8	Blood serum/plasma	Detection of IgM to hepatitis A virus
0356	HAV-antigen-EIA-BEST	12×8	Feces	Detection of hepatitis A virus antigen
0362 <b>CE</b>	Vectohep A-IgG	12×8	Blood serum/plasma, blood products	Detection and quantification of IgG to hepatitis A virus
<b>Hepatitis B</b>				
0543	HBsAg-EIA-BEST Set 2	24×8	Blood serum/plasma	Detection of HBsAg
0544	HBsAg-EIA-BEST Set 3	12×8	Blood serum/plasma	Two-step analysis
0546	HBsAg-confirmed-EIA-BEST	6×8	Blood serum/plasma	Sensitivity: 0.01 IU/ml
0557	Vectohep B-HBs-antigen Set 2	24×8	Blood serum/plasma	Confirmation of the presence of HBsAg
0556	Vectohep B-HBs-antigen Set 3	12×8	Blood serum/plasma	Sensitivity: 0.01 IU/ml
0559	Vectohep B-HBs-antigen Set 4	60×8	Blood serum/plasma	One-step analysis
0571	Vectohep HBsAg plus Set 1	12×8	Blood serum/plasma	Sensitivity: 0.05/0.01 IU/ml
0572	Vectohep HBsAg plus Set 2	24×8	Blood serum/plasma	Ready-to-use reagents
0573	Vectohep HBsAg plus Set 3	60×8	Blood serum/plasma	Sensitivity: 0.05/0.01 IU/ml
0558	Vectohep B-HBs-antigen- confirmatory test Set 1	6×8	Blood serum/plasma	Confirmation of the presence of HBsAg
				One-step analysis
				Sensitivity: 0.05/0.01 IU/ml

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b>Hepatitis B</b>				
0545	<b>HBsAg-EIA-BEST-quantitative</b>	12×8	Blood serum/plasma	Quantification of HBsAg Sensitivity: 0.05 IU/ml Measurement range: 0–10 IU/ml
0562	<b>VectoHBsAg-antibodies</b>	12×8	Blood serum/plasma	Detection and quantification of antibodies to HBsAg
0564	<b>VectoHBcAg-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to hepatitis B core antigen
0566	<b>VectoHBcAg-antibodies</b>	12×8	Blood serum/plasma	Detection of total antibodies to hepatitis B core antigen
0574	<b>HepaBest anti-HBc-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to hepatitis B core antigen
0576	<b>VectoHBe-antigen</b>	12×8	Blood serum/plasma	Detection of E antigen of hepatitis B virus
0578	<b>VectoHBe-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to HBe antigen of hepatitis B virus
<b>Hepatitis C</b>				
0761	<b>Best anti-HCV-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to hepatitis C virus
<b>Best anti-HCV</b>				
0772	Set 1	12×8	Blood serum/plasma	Detection of IgG and IgM to hepatitis C virus
0773	Set 2	24×8	Blood serum/plasma	
0775	Set 3	60×8	Blood serum/plasma	
0774	<b>Best anti-HCV-spectrum</b>	6×4	Blood serum/plasma	Detection of IgG and IgM to the antigens of hepatitis C virus (core, NS3, NS4, NS5)
0776	<b>Best anti-HCV-confirmatory test</b>	12×4	Blood serum/plasma	Confirmation of the presence of IgG and IgM to hepatitis C virus
<b>HCV Ag/Ab-EIA-BEST</b>				
0777	Set 1	12×8	Blood serum/plasma	Detection of antibodies to hepatitis C virus and HCV core antigen
0778	Set 2	24×8	Blood serum/plasma	
0779	<b>HCV core-antigen-EIA-BEST</b>	12×8	Blood serum/plasma	Detection of hepatitis C virus core antigen
<b>Hepatitis D</b>				
0952	<b>Vectohep D-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to hepatitis D virus
0954	<b>Vectohep D-antibodies</b>	12×8	Blood serum/plasma	Detection of total antibodies to hepatitis D virus

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
<b>Hepatitis E</b>				
1056	<b>Vectorhep E-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to hepatitis E virus
1058	<b>Vectorhep E-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to hepatitis E virus
<b>Sexually transmitted infections (STIs)</b>				
<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
<b>Syphilis</b>				
1852	<b>RecombiBest antipallidum-IgG</b> Set 2	12×8	Blood serum/plasma	Detection of IgG to <i>Treponema pallidum</i>
1856	<b>RecombiBest antipallidum-total antibodies</b> Set 2	12×8	Blood serum/plasma	Detection of total antibodies to <i>Treponema pallidum</i>
1857	Set 3	24×8		
1875	Set 4	60×8		
1858	<b>RecombiBest antipallidum-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to <i>Treponema pallidum</i>
1860	<b>Treponema pallidum-blot-BEST</b>	24	Blood serum/plasma, cerebrospinal fluid	Confirmation of the presence of antibodies to <i>Treponema pallidum</i> by immunoblot
1812	<b>LumiBest antipallidum</b>	80	Blood serum, cerebrospinal fluid	Detection of antibodies to <i>Treponema pallidum</i> by immunofluorescence
1820	<b>TPHA-BEST</b>	100	Blood serum, cerebrospinal fluid	Detection of antibodies to <i>Treponema pallidum</i> by hemagglutination assay
1822	<b>Anticardiolipin-RPR-BEST</b>	400	Blood serum/plasma, cerebrospinal fluid	Detection of syphilis-associated antibodies to cardiolipin by the precipitation reaction
<b>Chlamydiosis</b>				
1964	<b>ChlamyBest C.trachomatis-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to <i>Chlamydia trachomatis</i>
1966	<b>ChlamyBest C.trachomatis-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to <i>Chlamydia trachomatis</i>
1968	<b>ChlamyBest C.trachomatis-IgA</b>	12×8	Blood serum/plasma	Detection of IgA to <i>Chlamydia trachomatis</i>
1972	<b>ChlamyBest cHSP60-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to <i>Chlamydia trachomatis</i> heat shock protein (cHSP60)
1982	<b>ChlamyBest MOMP+ Pgp3-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to <i>Chlamydia trachomatis</i> major outer membrane protein (MOMP) and plasmic encoded protein (Pgp3)

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
<b>Trichomoniasis</b>				
2052	Trichomonas vaginalis-IgG-EIA-BEST	12x8	Blood serum/plasma	Detection of IgG to <i>Trichomonas vaginalis</i>
2058	Trichomonas vaginalis-IgA-EIA-BEST	12x8	Blood serum/plasma	Detection of IgA to <i>Trichomonas vaginalis</i>
<b>Ureaplasmosis</b>				
2254	Ureaplasma urealyticum-IgG-EIA-BEST	12x8	Blood serum/plasma	Detection of IgG to <i>Ureaplasma urealyticum</i> antigens
2258	Ureaplasma urealyticum-IgA-EIA-BEST	12x8	Blood serum/plasma	Detection of IgG to <i>Ureaplasma urealyticum</i> antigens
<b>Mycoplasmosis</b>				
4352	Mycoplasma hominis-IgG-EIA-BEST	12x8	Blood serum/plasma	Detection of IgG to <i>Mycoplasma hominis</i>
4358	Mycoplasma hominis-IgA-EIA-BEST	12x8	Blood serum/plasma	Detection of IgA to <i>Mycoplasma hominis</i>
<b>Human herpes viruses and TORCH-infections</b>				
<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
<b>Toxoplasmosis</b>				
1752	VectoToxo-IgG	12x8	Blood serum/plasma	Detection and quantification of IgG to <i>Toxoplasma gondii</i>
1754	VectoToxo-IgA	12x8	Blood serum/plasma	Detection of IgA to <i>Toxoplasma gondii</i>
1756	VectoToxo-IgM	12x8	Blood serum/plasma	Detection of IgM to <i>Toxoplasma gondii</i>
1760	Toxo-IgM-EIA-BEST	12x8	Blood serum/plasma	Detection of IgM to <i>Toxoplasma gondii</i> using the antibody capture method
1762	VectoToxo-IgG-avidity	6x8	Blood serum/plasma	Determination of the avidity index of IgG to <i>Toxoplasma gondii</i>
1764	VectoToxo-antibodies	12x8	Blood serum/plasma	Detection of total antibodies to <i>Toxoplasma gondii</i>
<b>Rubella</b>				
2552	VectoRubella-IgG	12x8	Blood serum/plasma	Detection and quantification of IgG to rubella virus
2554	VectoRubella-IgM	12x8	Blood serum/plasma	Detection of IgM to rubella virus
2560	Rubella-IgM-EIA-BEST	12x8	Blood serum/plasma	Detection of IgM to rubella virus using the antibody capture method
2556	VectoRubella-IgG-Avidity	6x8	Blood serum/plasmas	Determination of the avidity index of IgG to rubella virus

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b>Cytomegalovirus infection</b>				
1552	<b>VectoCMV-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to cytomegalovirus
1554	<b>VectoCMV-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to cytomegalovirus
1556	<b>CMV-IgG-EIA-BEST</b>	12×8	Blood serum/plasma	Detection and quantification of IgG to cytomegalovirus
1557	<b>CMV-IgM-EIA-BEST</b>	12×8	Blood serum/plasma	Detection of IgM to cytomegalovirus using the antibody capture method
1558	<b>VectoCMV-IgG-avidity</b>	6×8	Blood serum/plasma	Determination of the avidity index of IgG to cytomegalovirus
1560	<b>CMV-IgG-blot-BEST</b>	20	Blood serum/plasma	Confirmation of the presence of IgG to cytomegalovirus by immunoblot
1566	<b>VectoCMV-IEA-antibodies</b>	12×8	Blood serum/plasma	Detection of IgG and IgM to cytomegalovirus immediate early antigen
<b>Herpes simplex virus types 1 and 2 infection</b>				
2152	<b>VectoHSV-1,2-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to herpes simplex virus types 1 and 2
2154	<b>VectoHSV-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to herpes simplex virus types 1 and 2
2156	<b>VectoHSV-1,2-IgG-avidity</b>	6×8	Blood serum/plasma	Determination of the avidity index of IgG to herpes simplex virus types 1 and 2
2158	<b>VectoHSV-1-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to herpes simplex virus type 1
2180	<b>VectoHSV-2-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to herpes simplex virus type 2
2181	<b>VectoHSV-2-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to herpes simplex virus type 2
2182	<b>VectoHSV-2-IgG-avidity</b>	6×8	Blood serum/plasma	Determination of the avidity index of IgG to herpes simplex virus type 2
<b>Epstein-Barr virus infection</b>				
2170 <b>CE</b>	<b>VectoEBV-NA-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to Epstein-Barr virus nuclear antigens
2172 <b>CE</b>	<b>VectoEBV-EA-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to Epstein-Barr virus early antigens
2176 <b>CE</b>	<b>VectoEBV-VCA-IgM</b>	12×8	Blood serum/plasma	Detection of IgM to Epstein-Barr virus viral capsid antigens
2183	<b>VectoEBV-VCA-IgG-avidity</b>	6×8	Blood serum/plasma	Determination of the avidity index of IgG to Epstein-Barr virus viral capsid antigens
2184	<b>VectoEBV-VCA-IgG</b>	12×8	Blood serum/plasma	Detection of IgG to Epstein-Barr virus viral capsid antigens

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b>Human herpes virus type 6 infection</b>				
2166	HHV-6-IgG-EIA-BEST	12x8	Blood serum/plasma	Detection of IgG to human herpes virus type 6
2167	HHV-6-IgM-EIA-BEST	12x8	Blood serum/plasma	Detection of IgM to human herpes virus type 6 using the antibody capture method
<b>Human herpes virus type 8 infection</b>				
2160	VectoHHV-8-IgG	12x8	Blood serum/plasma	Detection of IgG to human herpes virus type 8
<b>Varicella-zoster virus infection</b>				
2186	VectoVZV-gE-IgG	12x8	Blood serum/plasma	Detection of IgG to varicella-zoster virus glycoprotein E
2188	VectoVZV-IgM	12x8	Blood serum/plasma	Detection of IgM to varicella-zoster virus
2192	VectoVZV-IgG	12x8	Blood serum/plasma	Detection of IgG to varicella-zoster virus
<b>Mycoses</b>				
Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b>Candidiasis</b>				
4652	Candida-IgG-EIA-BEST	12x8	Blood serum/plasma	Detection of IgG to the fungi of <i>Candida</i> genus
4654	Candida-IgM-EIA-BEST	12x8	Blood serum/plasma	Detection of IgM to the fungi of <i>Candida</i> genus
4656	Candida-IgA-EIA-BEST	12x8	Blood serum/plasma	Detection of IgA to the fungi of <i>Candida</i> genus
<b>Aspergillosis</b>				
4752	Aspergillus-IgG-EIA-BEST	12x8	Blood serum/plasma	Detection of IgG to <i>Aspergillus fumigatus</i> antigens
<b>Tick-borne and zoonotic infections</b>				
Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b>Tick-borne encephalitis</b>				
1152 	VectoTBEV-IgM	12x8	Blood serum/plasma	Detection of IgM to tick-borne encephalitis virus
1154	VectoTBEV-antigen	12x8	Ticks, cerebrospinal fluid	Detection of tick-borne encephalitis virus antigen
1156 	VectoTBEV-IgG	12x8	Blood serum/plasma	Detection and quantification of IgG to tick-borne encephalitis virus

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b>Borreliosis</b>				
1452 CE	LymeBest-IgG	12×8	Blood serum/plasma	Detection of IgG to the causative agent of Lyme disease
1454 CE	LymeBest-IgM	12×8	Blood serum/plasma	Detection of IgM to the causative agent of Lyme disease
<b>Hemorrhagic fever with renal syndrome</b>				
4902 CE	VectoHanta-IgG	12×8	Blood serum/plasma	Detection of IgG to hantaviruses
4903	Hanta-IgG-express-BEST	10	Blood serum/plasma, whole blood	Detection of IgG to hantaviruses by immunochromatography
4904 CE	VectoHanta-IgM	12×8	Blood serum/plasma	Detection of IgM to hantaviruses
4905	Hanta-IgM-express-BEST	10	Blood serum/plasma, whole blood	Detection of IgM to hantaviruses by immunochromatography
<b>Crimean-Congo hemorrhagic fever</b>				
5052 CE	VectoCrimean-CHF-IgG	12×8	Blood serum/plasma	Detection of IgG to Crimean-Congo hemorrhagic fever virus
5054 CE	VectoCrimean-CHF-IgM	12×8	Blood serum/plasma	Detection of IgM to Crimean-Congo hemorrhagic fever virus
5056 CE	VectoCrimean-CHF-antigen	12×8	Blood serum/plasma	Detection of Crimean-Congo hemorrhagic fever virus antigen
<b>West Nile fever</b>				
5150 CE	VectoNile-IgM	12×8	Blood serum/plasma	Detection of IgM to West Nile virus
5152 CE	VectoNile-IgG	12×8	Blood serum/plasma	Detection of IgG to West Nile virus
5154 CE	VectoNile-IgG-avidity	6×8	Blood serum/plasma	Determination of the avidity index of IgG to West Nile virus
<b>Yersiniosis</b>				
3202 CE	Yersinia-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to the causative agents of yersiniosis
3204 CE	Yersinia-IgA-EIA-BEST	12×8	Blood serum/plasma	Detection of IgA to the causative agents of yersiniosis
3206 CE	Yersinia-IgM-EIA-BEST	12×8	Blood serum/plasma	Detection of IgM to the causative agents of yersiniosis

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b>Brucellosis</b>				
3652	Brucellosis-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to the causative agent of brucellosis
3654	Brucellosis-IgA-EIA-BEST	12×8	Blood serum/plasma	Detection of IgA to the causative agent of brucellosis
3656	Brucellosis-IgM-EIA-BEST	12×8	Blood serum/plasma	Detection of IgM to the causative agent of brucellosis
3658	Brucellosis-antibodies-EIA-BEST	12×8	Blood serum/plasma	Detection of total antibodies to the causative agent of brucellosis

### Vaccine-preventable diseases

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b>Measles</b>				
1356 <b>CE</b>	VectoMeasles-IgG	12×8	Blood serum/plasma	Detection and quantification of IgG to measles virus
1358 <b>CE</b>	VectoMeasles-IgM	12×8	Blood serum/plasma	Detection of IgM to measles virus
<b>Mumps</b>				
2602 <b>CE</b>	VectoMumps-IgG	12×8	Blood serum/plasma	Detection of IgG to mumps virus
2606	Mumps-IgM-EIA-BEST	12×8	Blood serum/plasma	Detection of IgM to mumps virus using the antibody capture method

### Respiratory infections

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b><i>Chlamydophila pneumoniae</i> infection</b>				
1944	Chlamydophila pneumoniae-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to <i>Chlamydophila pneumoniae</i>
1946	Chlamydophila pneumoniae-IgM-EIA-BEST	12×8	Blood serum/plasma	Detection of IgM to <i>Chlamydophila pneumoniae</i>
1948	Chlamydophila pneumoniae-IgA-EIA-BEST	12×8	Blood serum/plasma	Detection of IgA to <i>Chlamydophila pneumoniae</i>
<b>Respiratory mycoplasmosis</b>				
4362	Mycoplasma pneumoniae-IgG-EIA-BEST	12×8	Blood serum	Detection of IgG to <i>Mycoplasma pneumoniae</i>
4366	Mycoplasma pneumoniae-IgM-EIA-BEST	12×8	Blood serum	Detection of IgM to <i>Mycoplasma pneumoniae</i>
4368	Mycoplasma pneumoniae-IgA-EIA-BEST	12×8	Blood serum	Detection of IgA to <i>Mycoplasma pneumoniae</i>

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
<b>Coronavirus infection</b>				
5501 <b>CE</b>	SARS-CoV-2-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to SARS-CoV-2
5502 <b>CE</b>	SARS-CoV-2-IgM-EIA-BEST	12×8	Blood serum/plasma	Detection of IgM to SARS-CoV-2
5503	SARS-CoV-2-IgA-EIA-BEST	12×8	Blood serum/plasma	Detection of IgA to SARS-CoV-2
5504	SARS-CoV-2-Ab total-EIA-BEST	12×8	Blood serum/plasma	Detection of total antibodies to SARS-CoV-2
5505 <b>CE</b>	SARS-CoV-2-IgG quantitative-EIA-BEST	12×8	Blood serum/plasma	Quantification of IgG to SARS-CoV-2 Sensitivity: 0.5 BAU/ ml
5508	SARS-CoV-2-Ag-express-BEST	25	Nasal and oropharyngeal swabs	Detection of SARS-CoV-2 nucleocapsid antigen by immunochromatography
<b>Tuberculosis</b>				
Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
2352	AB-Tub-BEST	12×8	Blood serum/plasma	Detection of total antibodies to <i>Mycobacterium tuberculosis</i>
<b>Parasitic infections</b>				
Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
2752 <b>CE</b>	Toxocara-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to <i>Toxocara</i> antigens
2952	Opisthorchis-IgG-EIA-BEST (Set 1)	12×8	Blood serum/plasma	Detection of IgG to <i>Opisthorchis</i> antigens
2954	Opisthorchis-IgM-EIA-BEST	12×8	Blood serum/plasma	Detection of IgM to <i>Opisthorchis</i> antigens
2956	Opisthorchis-CIC-EIA-BEST	12×8	Blood serum/plasma	Detection of specific circulating immune complexes containing <i>Opisthorchis</i> antigens
2958	Clonorchis-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to <i>Clonorchis sinensis</i> antigens
3152 <b>CE</b>	Trichinella-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to <i>Trichinella</i> antigens
3154	Trichinella-IgM-EIA-BEST	12×8	Blood serum/plasma	Detection of IgM to <i>Trichinella</i> antigens
3356 <b>CE</b>	Echinococcus-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to <i>Echinococcus granulosus</i> antigens
3452 <b>CE</b>	Ascaris-IgG-EIA-BEST (Set 1)	12×8	Blood serum/plasma	Detection of IgG to <i>Ascaris lumbricoides</i> antigens

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
3454	Anisakis-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to the antigens of the nematodes of <i>Anisakis</i> genus
3456	Cysticercus-IgG-EIA-BEST	12×8	Blood serum/plasma	Detection of IgG to <i>Taenia solium</i> antigens
3552	Lamblia-antibodies-EIA-BEST (Set 1)	12×8	Blood serum/plasma	Detection of IgA, IgM, and IgG to <i>Giardia lamblia</i> antigens
3554	Lamblia-IgM-EIA-BEST (Set 1)	12×8	Blood serum/plasma	Detection of IgM to <i>Giardia lamblia</i> antigens
3556	Lamblia-antigen-EIA-BEST	12×8	Feces	Detection of <i>Giardia lamblia</i> antigens
3354	Helminths-IgG-EIA-BEST	24×8	Blood serum/plasma	Detection of IgG to the antigens of <i>Opisthorchis</i> , <i>Trichinella</i> , <i>Toxocara</i> , and <i>Echinococcus</i>

### Gastrointestinal diseases

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
<b>Helicobacter pylori infection</b>				
3752	Helicobacter pylori-CagA-antibodies-EIA-BEST	12×8	Blood serum/plasma	Detection of total antibodies to <i>Helicobacter pylori</i> CagA antigen
<b>Rotavirus infection</b>				
1652 <b>CE</b>	Rotavirus-antigen-EIA-BEST	12×8	Virus-containing culture fluids, feces, samples of drinking and waste water, water from open reservoirs	Detection of human rotavirus antigen
<b>Adenovirus infection</b>				
1654 <b>CE</b>	Adenovirus-antigen-EIA-BEST	12×8	Feces	Detection of human adenovirus antigen
<b>Norovirus infection</b>				
1656	Norovirus-antigen-EIA-BEST	12×8	Feces	Detection of noroviruses of genogroups I and II

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
<b>Celiac disease</b>				
3754	IgG-Gliadin-EIA-BEST	12×8	Blood serum	Quantification of anti-gliadin antibodies IgG Sensitivity: 1 U/ml Measurement range: 0–100 U/ml
3756	IgA-Gliadin-EIA-BEST	12×8	Blood serum	Quantification of anti-gliadin antibodies IgA Sensitivity: 1 U/ml Measurement range: 0–100 U/ml
3758 <b>CE</b>	IgA-Transglutaminase-EIA-BEST	12×8	Blood serum/plasma	Quantification of IgA autoantibodies to tissue transglutaminase Sensitivity: 2.5 U/ml Measurement range: 0–100 U/ml
3760 <b>CE</b>	IgG-Transglutaminase-EIA-BEST	12×8	Blood serum/plasma	Quantification of IgG autoantibodies to tissue transglutaminase Sensitivity: 2.5 U/ml Measurement range: 0–100 U/ml
<b>Atrophic gastritis</b>				
3762 <b>CE</b>	Pepsinogen 1-EIA-BEST	12×8	Blood serum	Quantification of pepsinogen 1 Sensitivity: 0.5 µg/l Measurement range: 0–200 µg/l
3764 <b>CE</b>	Pepsinogen 2-EIA-BEST	12×8	Blood serum	Quantification of pepsinogen 2 Sensitivity: 0.5 µg/l Measurement range: 0–50 µg/l
<b>Hormones</b>				
<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
<b>Thyroid function</b>				
3952	TSH-EIA-BEST	12×8	Blood serum/plasma	Quantification of thyroid stimulating hormone (TSH) Sensitivity: 0.05 mIU/l Measurement range: 0–16 mIU/l
3954	T3 total-EIA-BEST	12×8	Blood serum	Quantification of total triiodothyronine (T3) Sensitivity: 0.2 nmol/l Measurement range: 0–9 nmol/l
3956	T4 total-EIA-BEST	12×8	Blood serum	Quantification of total thyroxine (T4) Sensitivity: 5 nmol/l Measurement range: 0–400 nmol/l
3962	T4 free-EIA-BEST	12×8	Blood serum	Quantification of free thyroxine (T4) Sensitivity: 0.5 pmol/l Measurement range: 0–80 pmol/l

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
3970	T3 free-EIA-BEST	12×8	Blood serum	Quantification of free triiodothyronine (T3) Sensitivity: 0.5 pmol/l Measurement range: 0–20 pmol/l
3966	Tg-EIA-BEST	12×8	Blood serum	Quantification to thyroglobulin (Tg) Sensitivity: 1.0 ng/ml Measurement range: 0–300 ng/ml
3958	Anti-Tg-EIA-BEST	12×8	Blood serum	Quantification of total antibodies to thyroglobulin (Tg) Sensitivity: 5 IU/ml Measurement range: 0–2000 IU/ml
3968	Anti-TPO-EIA-BEST	12×8	Blood serum	Quantification of total antibodies to thyroid peroxidase (TPO) Sensitivity: 2 U/ml Measurement range: 0–1000 U/ml
<b>Reproductive system function</b>				
3980	DHEA-S-EIA-BEST	12×8	Blood serum/plasma	Quantification of DHEA-S Sensitivity: 0.02 µg/ml Measurement range: 0–10 µg/ml
3960	Prolactin-EIA-BEST	12×8	Blood serum	Quantification of prolactin Sensitivity: 15 mIU/l Measurement range: 0–3000 mIU/l
3972	Testosterone-EIA-BEST	12×8	Blood serum	Quantification of testosterone Sensitivity: 0.2 nmol/l Measurement range: 0–60 nmol/l
3974	FSH-EIA-BEST	12×8	Blood serum	Quantification of follicle-stimulating hormone (FSH) Sensitivity: 0.3 mIU/ml Measurement range: 0–100 mIU/ml
3976	LH-EIA-BEST	12×8	Blood serum	Quantification of luteinizing hormone (LH) Sensitivity: 0.3 mIU/ml Measurement range: 0–100 mIU/ml
3978	Progesterone-EIA-BEST	12×8	Blood serum	Quantification of progesterone Sensitivity: 0.4 nmol/l Measurement range: 0–100 nmol/l
3982	17-OH-progesterone-EIA-BEST	12×8	Blood serum/plasma	Quantification of 17-OH-progesterone Sensitivity: 0.11 nmol/l Measurement range: 0–100 nmol/l
<b>Pituitary-adrenal system function</b>				
3964	Cortisol-EIA-BEST	12×8	Blood serum	Quantification of cortisol Sensitivity: 5 nmol/l Measurement range: 0–1200 nmol/l
3980	DHEA-S-EIA-BEST	12×8	Blood serum/plasma	Quantification of DHEA-S Sensitivity: 0.02 µg/ml Measurement range: 0–10 µg/ml
3982	17-OH-progesterone-EIA-BEST	12×8	Blood serum/plasma	Quantification of 17-OH-progesterone Sensitivity: 0.11 nmol/l Measurement range: 0–100 nmol/l

## Diabetes

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
4001	C-peptide-EIA-BEST	12×8	Blood serum/plasma	Quantification of C-peptide Sensitivity: 15 pmol/l Measurement range: 0–5000 pmol/l
4002	Insulin-EIA-BEST	12×8	Blood serum/plasma	Quantification of insulin Sensitivity: 0.75 mIU/l Measurement range: 0–200 mIU/l

## Pregnancy

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
4154	HCG-EIA-BEST	12×8	Blood serum	Quantification of human chorionic gonadotropin (hCG) Sensitivity: 2 IU/l Measurement range: 0–500 IU/l
4158	free beta-HCG-EIA-BEST	12×8	Blood serum	Quantification of free beta-subunit of human chorionic gonadotropin (hCG) Sensitivity: 0.5 ng/ml Measurement range: 0–200 ng/ml
4160	PAPP-A-EIA-BEST	12×8	Blood serum	Quantification of pregnancy-associated plasma protein A (PAPP-A) Sensitivity: 20 mIU/l Measurement range: 0–10000 mIU/l
3960	Prolactin-EIA-BEST	12×8	Blood serum	Quantification of prolactin Sensitivity: 15 mIU/l Measurement range: 0–3000 mIU/l
8456	AFP-EIA-BEST	12×8	Blood serum	Quantification of alpha-fetoprotein (AFP) Sensitivity: 2.5 IU/ml Measurement range: 0–400 IU/ml
8468	PSG1-EIA-BEST	12×8	Blood serum	Quantification of pregnancy-specific beta-1-glycoprotein 1 Sensitivity: 1.0 ng/ml Measurement range: 0–200 ng/ml
8552	Ferritin-EIA-BEST	12×8	Blood serum/plasma	Quantification of ferritin Sensitivity: 2.5 ng/ml Measurement range: 0–500 ng/ml

## Anemia

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
8552	Ferritin-EIA-BEST	12×8	Blood serum/plasma	Quantification of ferritin Sensitivity: 2.5 ng/ml Measurement range: 0–500 ng/ml
8776	Erythropoietin-EIA-BEST	12×8	Blood serum/plasma	Quantification of erythropoietin Sensitivity: 0.5 mIU/ml Measurement range: 0–200 mIU/ml

## Tumor markers

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
4154	HCG-EIA-BEST	12×8	Blood serum	Quantification of human chorionic gonadotropin (hCG) Sensitivity: 2 IU/l Measurement range: 0–500 IU/l
4158	free beta-HCG-EIA-BEST	12×8	Blood serum	Quantification of free beta-subunit of human chorionic gonadotropin (hCG) Sensitivity: 0.5 ng/ml Measurement range: 0–200 ng/ml
8437	CYFRA 21-1-EIA-BEST Set 1	96	Blood serum/plasma	Quantification of soluble fragments of cytokeratin 19 Sensitivity: 0.15 ng/ml Measurement range: 0–50 ng/ml
8454	CEA-EIA-BEST	12×8	Blood serum	Quantification of carcinoembryonic antigen (CEA) Sensitivity: 3.3 mIU/ml Measurement range: 0–880 mIU/ml
8455	CA 72-4-EIA-BEST	12×8	Blood serum	Quantification of CA 72-4 Sensitivity: 0.16 U/ml Measurement range: 0–100 U/ml
8456	AFP-EIA-BEST	12×8	Blood serum	Quantification of alpha-fetoprotein (AFP) Sensitivity: 2.5 IU/ml Measurement range: 0–400 IU/ml
8453	PSA total-EIA-BEST variant 2	12×8	Blood serum	Quantification of total prostate-specific antigen (PSA) Sensitivity: 0.01 ng/ml Measurement range: 0–5 ng/ml
8460	PSA free-EIA-BEST	12×8	Blood serum	Quantification of free prostate-specific antigen (PSA) Sensitivity: 0.05 ng/ml Measurement range: 0–5 ng/ml
8466	CA-125-EIA-BEST	12×8	Blood serum	Quantification of tumor marker CA-125 Sensitivity: 1.5 U/ml Measurement range: 0–400 U/ml
8467	HE4-EIA-BEST	12×8	Blood serum	Quantification of tumor marker HE4 Sensitivity: 1.5 pmol/l Measurement range: 0–1000 pmol/l
8468	PSG1-EIA-BEST	12×8	Blood serum	Quantification of pregnancy-specific beta-1-glycoprotein 1 Sensitivity: 1.0 ng/ml Measurement range: 0–200 ng/ml
8470	CA 19-9-EIA-BEST	12×8	Blood serum	Quantification of CA 19-9 Sensitivity: 1 U/ml Measurement range: 0–300 U/ml
8472	CA 15-3-EIA-BEST	12×8	Blood serum	Quantification of tumor marker CA 15-3 Sensitivity: 0.5 U/ml Measurement range: 0–250 U/ml

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
8476	NSE-EIA-BEST	12×8	Blood serum	Quantification of neuron specific enolase (NSE) Sensitivity: 0.5 ng/ml Measurement range: 0–130 ng/ml
8552	Ferritin-EIA-BEST	12×8	Blood serum/plasma	Quantification of ferritin Sensitivity: 2.5 ng/ml Measurement range: 0–500 ng/ml
3762 <b>CE</b>	Pepsinogen 1-EIA-BEST	12×8	Blood serum	Quantification of pepsinogen 1 Sensitivity: 0.5 µg/l Measurement range: 0–200 µg/l
3764 <b>CE</b>	Pepsinogen 2-EIA-BEST	12×8	Blood serum	Quantification of pepsinogen 2 Sensitivity: 0.5 µg/l Measurement range: 0–50 µg/l

### Acute phase proteins

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
9002	CRP-EIA-BEST highly sensitive	12×8	Blood serum	Highly sensitive quantification of C-reactive protein (CRP) Sensitivity: 0.05 IU/l Measurement range: 0–10 IU/l
9004	Procalcitonin-EIA-BEST	12×8	Blood serum/plasma	Quantification of procalcitonin Sensitivity: 0.04 ng/ml Measurement range: 0–12.8 ng/ml

### Thrombosis

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
9120	D-dimer-EIA-BEST	12×8	Blood plasma	Quantification of D-dimer Sensitivity: 10 ng/ml Measurement range: 0–3000 ng/ml

### Cardiac markers

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
9002	CRP-EIA-BEST highly sensitive	12×8	Blood serum	Highly sensitive quantification of C-reactive protein (CRP) Sensitivity: 0.05 IU/l Measurement range: 0–10 IU/l
9102	NTproBNP-EIA-BEST	12×8	Blood serum	Quantification of N-terminal prohormone of brain natriuretic peptide (NTproBNP) Sensitivity: 20 pg/ml Measurement range: 0–2000 pg/ml
9104	FABP-EIA-BEST	12×8	Blood serum	Quantification of fatty-acid-binding protein (FABP) Sensitivity: 0.05 ng/ml Measurement range: 0–15 ng/ml

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
9106	Troponin I-EIA-BEST	12×8	Blood serum	Quantification of troponin I Sensitivity: 0.02 ng/ml Measurement range: 0–6 ng/ml
9108	Myoglobin-EIA-BEST	12×8	Blood serum	Quantification of myoglobin Sensitivity: 4 ng/ml Measurement range: 0–1000 ng/ml
9120	D-dimer-EIA-BEST	12×8	Blood plasma	Quantification of D-dimer Sensitivity: 10 ng/ml Measurement range: 0–3000 ng/ml
9130	Cystatin C-EIA-BEST	12×8	Blood serum	Quantification of cystatin C Sensitivity: 0.05 µg/ml Measurement range: 0–10 µg/ml

### Renal function

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
9130	Cystatin C-EIA-BEST	12×8	Blood serum	Quantification of cystatin C Sensitivity: 0.05 µg/ml Measurement range: 0–10 µg/ml
8762	Interleukin-8-EIA-BEST	12×8	Blood serum, urine	Quantification of interleukin-8 Sensitivity: 2 pg/ml Measurement range: 0–250 pg/ml
8766	Interleukin-1 beta-EIA-BEST	12×8	Blood serum, urine	Quantification of interleukin-1 beta Sensitivity: 1 pg/ml Measurement range: 0–250 pg/ml
8768 <b>CE</b>	Interleukin-6-EIA-BEST	12×8	Blood serum, urine	Quantification of interleukin-6 Sensitivity: 0.5 pg/ml Measurement range: 0–300 pg/ml

### Systemic and autoimmune diseases

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
3754	IgG-Gliadin-EIA-BEST	12×8	Blood serum	Quantification of anti-gliadin antibodies IgG Sensitivity: 1 U/ml Measurement range: 0–100 U/ml
3756	IgA-Gliadin-EIA-BEST	12×8	Blood serum	Quantification of anti-gliadin antibodies IgA Sensitivity: 1 U/ml Measurement range: 0–100 U/ml
3758 <b>CE</b>	IgA-Transglutaminase-EIA-BEST	12×8	Blood serum/plasma	Quantification of IgA autoantibodies to tissue transglutaminase Sensitivity: 2.5 U/ml Measurement range: 0–100 U/ml
3760 <b>CE</b>	IgG-Transglutaminase-EIA-BEST	12×8	Blood serum/plasma	Quantification of IgG autoantibodies to tissue transglutaminase Sensitivity: 2.5 U/ml Measurement range: 0–100 U/ml

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
3958	Anti-Tg-EIA-BEST	12×8	Blood serum	Quantification of total antibodies to thyroglobulin (Tg) Sensitivity: 5 IU/ml Measurement range: 0–2000 IU/ml
3968	Anti-TPO-EIA-BEST	12×8	Blood serum	Quantification of total antibodies to thyroid peroxidase (TPO) Sensitivity: 2 U/ml Measurement range: 0–1000 U/ml
8652	RF-IgM-EIA-BEST	12×8	Blood serum	Quantification of M-class rheumatoid factor Sensitivity: 1.0 U/ml Measurement range: 0–300 U/ml
8654	RF-total-EIA-BEST	12×8	Blood serum	Quantification of total rheumatoid factor Sensitivity: 1.0 U/ml Measurement range: 0–300 U/ml
8656	Vecto-dsDNA-IgG	12×8	Blood serum	Quantification of autoimmune IgG to double-stranded DNA Sensitivity: 2 IU/ml Measurement range: 0–200 IU/ml
8658	Vecto-ssDNA-IgG	12×8	Blood serum	Quantification of autoimmune IgG to single-stranded DNA Sensitivity: 2 U/ml Measurement range: 0–200 U/ml
8660	IgE total-EIA-BEST	12×8	Blood serum	Quantification of total IgE Sensitivity: 2.5 IU/ml Measurement range: 0–800 IU/ml

### Humoral immune status

<i>Cat. No.</i>	<i>Kit name</i>	<i>Number of tests</i>	<i>Clinical specimens</i>	<i>Comments</i>
8660	IgE total-EIA-BEST	12×8	Blood serum	Quantification of total IgE Sensitivity: 2.5 IU/ml Measurement range: 0–800 IU/ml
8662	IgG total-EIA-BEST	12×8	Blood serum	Quantification of total IgG Sensitivity: 2.5 U/ml Measurement range: 0–300 U/ml
8664	IgM total-EIA-BEST	12×8	Blood serum	Quantification of total IgM Sensitivity: 4 U/ml Measurement range: 0–400 U/ml
8666	IgA total-EIA-BEST	12×8	Blood serum	Quantification of total IgA Sensitivity: 1.5 U/ml Measurement range: 0–300 U/ml
8674	Immunoscreen-G, M, A-EIA-BEST	4×8	Blood serum	Quantification of total IgA, IgM, IgG
8668	IgA secretory-EIA-BEST	12×8	Blood serum	Quantification of secretory IgA (sIgA) Sensitivity: 0.35 mg/l Measurement range: 0–20 mg/l

## Cytokines

Cat. No.	Kit name	Number of tests	Clinical specimens	Comments
8752	gamma-Interferon-EIA-BEST	12×8	Blood serum	Quantification of gamma-interferon Sensitivity: 2 pg/ml Measurement range: 0–1000 pg/ml
8754	Interleukin-4-EIA-BEST	12×8	Blood serum	Quantification of interleukin-4 Sensitivity: 0.4 pg/ml Measurement range: 0–100 pg/ml
8756	alpha-TNF-EIA-BEST	12×8	Blood serum	Quantification of tumor necrosis factor alpha Sensitivity: 1 pg/ml Measurement range: 0–250 pg/ml
8758	alpha-Interferon-EIA-BEST	12×8	Blood serum	Quantification of interferon alpha Sensitivity: 5 pg/ml Measurement range: 0–500 pg/ml
8760	alpha-Interferon-autoimmune antibodies-EIA-BEST	12×8	Blood serum	Quantification of autoimmune antibodies to interferon alpha Sensitivity: 0.4 ng/ml Measurement range: 0–100 ng/ml
8762	Interleukin-8-EIA-BEST	12×8	Blood serum, urine	Quantification of interleukin-8 Sensitivity: 2 pg/ml Measurement range: 0–250 pg/ml
8766	Interleukin-1 beta-EIA-BEST	12×8	Blood serum, urine	Quantification of interleukin-1 beta Sensitivity: 1 pg/ml Measurement range: 0–250 pg/ml
8768 <b>CE</b>	Interleukin-6-EIA-BEST	12×8	Blood serum, urine	Quantification of interleukin-6 Sensitivity: 0.5 pg/ml Measurement range: 0–300 pg/ml
8770	Interleukin-18-EIA-BEST	12×8	Blood serum	Quantification of interleukin-18 Sensitivity: 2 pg/ml Measurement range: 0–1000 pg/ml
8772	Interleukin-2-EIA-BEST	12×8	Blood serum	Quantification of interleukin-2 Sensitivity: 2 pg/ml Measurement range: 0–500 pg/ml
8774	Interleukin-10-EIA-BEST	12×8	Blood serum	Quantification of interleukin-10 Sensitivity: 1 pg/ml Measurement range: 0–500 pg/ml
8776	Erythropoietin-EIA-BEST	12×8	Blood serum/plasma	Quantification of erythropoietin Sensitivity: 0.5 mIU/ml Measurement range: 0–200 mIU/ml
8782	MCP-1-EIA-BEST	12×8	Blood serum	Quantification of monocyte chemoattractant protein-1 (MCP-1) Sensitivity: 15 pg/ml Measurement range: 0–2000 pg/ml
8784	VEGF-EIA-BEST	12×8	Blood serum	Quantification of vascular endothelial growth factor (VEGF) Sensitivity: 10 mIU/ml (pg/ml) Measurement range: 0–2000 mIU/ml (pg/ml)

